

1 10A NCAC 15 .0801 is proposed for amendment as follows:

2  
3 **SECTION .0800 REQUIREMENTS FOR ANALYTICAL X-RAY EQUIPMENT NON-HUMAN USE OF**  
4 **RADIATION GENERATING DEVICES**

5  
6 **10A NCAC 15 .0801 PURPOSE AND SCOPE**

7 ~~This Section provides special requirements for analytical x-ray equipment which are in addition to, and not in~~  
8 ~~substitution for, applicable requirements in the other sections of this Chapter.~~

9 (a) This Section provides special requirements for use of ionizing radiation generating devices (RGDs) operating  
10 above five thousand electron volts (5 keV), but below one million electron volts (1 MeV) that are in addition to, and  
11 not in substitution for, applicable requirements in the other Sections of this Chapter.

12 (b) This Section does not pertain to radiation safety requirements for x-ray equipment that is covered in other Sections  
13 of this Chapter (e.g., x-rays in the healing arts in Section .0600, and particle accelerators in Section .0900.)

14  
15 *History Note: Authority G.S. 104E-7;*  
16 *Eff. February 1, ~~1980~~ 1980;*  
17 *Amended Eff. October 1, 2015.*